

**DEPENDABILITY OF BODY AREA WIRELESS
SENSOR NETWORKS IN ASSISTIVE CARE LOOP
FRAMEWORK**

By

Venkatakrisnan Balasubramanian Appiah

A THESIS SUBMITTED FOR THE DEGREE OF

DOCTOR OF PHILOSOPHY

IN THE

FACULTY OF ENGINEERING AND INFORMATION TECHNOLOGY

UNIVERSITY OF TECHNOLOGY, SYDNEY

© 2012 Venkatakrishnan Balasubramanian Appiah

All Rights Reserved

CERTIFICATE OF AUTHORSHIP

I certify that the work in this thesis has not previously been submitted for a degree nor has it been submitted as part of requirements for a degree except as fully acknowledged within the text.

I also certify that the thesis has been written by me. Any help that I have received in my research work and the preparation of the thesis itself has been acknowledged. In addition, I certify that all information sources and literature used are indicated in the thesis.

Signed:

B. Uv

Venkatakrishnan Balasubramanian Appiah

Date:

02-NOV-2012

In the memory of my beloved *Balu Appa*

Acknowledgement

I thank my supervisor *Prof Doan B Hoang*, not only for his valuable guidance throughout my PhD tenure, but also for the care showed me during the hard times. In each and every meeting we had, with his immense research experience, he brought me encouragement and taught me how to find solution for the various research issues in different ways, which induced me to instigate new ideas into my research. His talks were not only focused towards research, but also on real life scenarios. This has helped me to complete my PhD in spite of the several hindrances, motivated me to handle difficult situations in real life and made me a better person. Without his help in terms of technicalities, funds and opportunity, I would not have reached this stage. I wish to express my gratitude to *Prof San Murugesan* for providing the opportunity to work under him as a Research Assistant in mobile application developments and for showing me the way to pursue my PhD. I thank my master research thesis supervisor, *Dr Selvakenedy Selvadurai*, for making me understand the doctrine of commitment that is essential for becoming a researcher.

I owe much to my *friend* and my *best man*, *Dr Venkateshan Balakrishnan* and his wife, *Mrs Shobana Venkateshan*, for their timely help during my ill health and their encouragement provided for finishing my PhD each time whenever I rang or visited them. I am grateful for all the time he spent in reviewing my research papers, his motivational talk and readiness to answer most of my queries regarding the research approach. I wholeheartedly thank them for the support they provided in the last few years of my tenure when I was going through a rough phase, both in my professional as well as in my personal life. I wish to thank *Dr Tanveer Zia* for collaborating with me and for helping me to obtain a casual academic role that helped me financially in many ways. Overall, he is one of the *trustworthy friend* and a 'go-to' man in case of any professional needs. I wish to thank and appreciate *Dr Priya* for collaborating with me in exploring the details of the medical diagnoses. As a physician, she helped me find the relevant medical research materials and enabled me understand how the basic medical diagnoses were made for diseases. I

am thankful and grateful to *Mr Tony Roberts* for editing my thesis at many short notices and satisfying my untimely requirements/demands that affected his social life. I thank my co-researcher and friend *Mrs Nor Faizah Ahmed* for providing me with the screenshots, templates and any professional help that I needed to finish my thesis. She made me so comfortable that I can disturb her anytime when I need anything regarding research. I wish to thank my student *Mr. Sarbodaya Kandel* for his contributed code base, upon my guidance.

I am indebted to my *first friend* in Australia and *mentor Mr Swaminathan Balasubramanian* for all his help he showered me without any expectation and a humble soul who is ready to lend an ear whenever I needed him, not only during my research tenure, but also all throughout these years. I am obliged to my friend *Mr Satya Ranjith Reddy* for his helping hand when I desperately needed one, during the times of hardships in my personal life. His help lay in taking care of my lectures/tutorials during my absence and as a friend, always ready to listen to my problems. I owe so much to my *brother, Mr Shankar Amaravelu*, for the time he spent during those long nights during the preparation of my final presentation/paper; he was readily available whenever I needed him and my 'go-to' man in case of any personal needs. I wish to thank *Mr Sathish Balachander* and his wife, *Mrs Shameena*, for taking care of me during my sudden hospitalization; they came from nowhere and helped me beyond the limits one could expect. I am obliged to *Mr Arunkumar Athimulam* and *Mr Ganesh Udayakumar* for helping me to settle down after my father's demise and their unconditional love towards me during the initial stage of my married life; a special thanks to *Mr Ganesh Udayakumar* for taking care by ringing me every week to find out my thesis progress. I thank *Mr Ashok* and *Mr Harpreet Singh* for being there whenever I needed a break from my work. I thank my schoolmates, *Mr Vishak*, *Mr Prashanth*, *Mr Jayakanthan*, *Dr Karthikraj*, *Mr. Arunkumar* and *Mr Kalyan* for being with me throughout my life and helping both me and my parents during all the bad and good occasions during my tenure. I am privileged to have my GCT college mates and *forever friends*, *Mr Madhukar Nimmagadda* , *Mr Balaji Shanmugam*, *Mr AP*

Suresh Kumar and *Mr Balamurugan Kannan* with their constant encouragement and being readily available if I needed to discuss any technical issues during the application development. I am grateful to *Mr Balaji Shanmugam* and *Mr Balamurugan Kannan* for his unconditional love and the gestures he shows towards his friends. I am honestly proud to have their friendship.

I am indebted to my *sister Mrs Vijayarani Natarajan* for the love, encouragement and the support she showed me throughout my higher studies. From her I learned the commitment it takes to become a professional. I thank her for teaching me the mathematics that sowed the seed for my lasting career — I have come long way with that knowledge. Without her financial support, I would not have enjoyed freedom during my higher studies. I also indebted to my *elder brother, Dr Rajagopal Balasubramanian* for helping me financially during my higher studies, he taught me the patience so necessary to become a better researcher. His help in proofreading some of my research papers and in writing formal letters to professors are worth mentioning. I am blessed to have my beloved father, *Dr Appiah Balasubramanian* and my beloved mother, *Mrs Kalavathy Balasubramanian*, who taught me the value of life, the power of knowledge and the benefits of self-discipline. They were full of wisdom and patience, and spent all their lifetime building a family for their children. I wholeheartedly dedicate this thesis to their selfless attitude. To my wife, *Mrs Buvaneswari Venkatakrishnan*, for her unconditional love, affection and the tolerance she has showed all these years. During all those frustrating days in the office and sleepless nights while writing the thesis, she made sure that I had a sufficient supply of food and tea. I owe her for the patience she had shown all these years, in waiting for me to complete my research and lead a normal life. Her optimism and desire to help me throughout my tenure during all the weekends I spent doing the research motivated me to reach the finishing line. With all her smiles, she takes away all my frustration and stress. She made me believe that there is nothing I cannot accomplish with her by my side.

I owe the *Bhagavan* (The Absolute Truth) for manifesting my soul in this world with a purpose that might benefit, at least the slightest, for the welfare of humankind.

List of Award and Publications

Award

- 2008 **Best Collaborative Research Project** at UTS in **HEALTH RESEARCH SHOWCASE** for creating *Assistive Maternity Care* Application
Research Team: Caroline Homer, Doan Hoang, **Venki Balasubramanian**, Nor Faizah Ahmad, Nicky Leap, Maralyn Foureur and Elaine Lawrence.

Journal

- **V. Balasubramanian**, and D. B. Hoang, "Reliability Measure Model for Assistive Care Loop Framework using Wireless Sensor Networks." *International Journal of Healthcare Engineering*, vol. 1, no.2, pp. 239-254, 2010.
- C.S Homer, C.J. Catling-Paull, D. Sinclair, N. Ahmad, **V. Balasubramanian**, M. Foureur, D.B. Hoang and E.M. Lawrence, "Developing an interactive electronic maternity record", *British Journal of Midwifery*, vol. 18, no. 6, pp. 384-389, 2010.

Conference Proceedings

- **V. Balasubramanian**, T. A. Zia and Doan. B. Hoang, "Addressing the Confidentiality and Integrity of Assistive Care Loop Framework using Wireless Sensor Networks", in *Proceedings of the 21st International Conference on System Engineering*, Las Vegas, NV, USA, 2011, pp. 416-421.
- **V. Balasubramanian**, and D. B. Hoang, "Availability Measure Model for Assistive Care Loop Framework using Wireless Sensor Networks", in *Proceedings of the 6th International Conference on Intelligent Sensors, Sensor Networks and Information Processing*, Brisbane, Australia, 2010, pp. 281-286.
- **V. Balasubramanian** and D. B. Hoang, "SOAP-based Assistive Care Loop using Wireless Sensor Networks", in *Proceedings of the 1st IEEE International*

Symposium on IT in Medicine and Education, IEEE Computer Society, Xiamen, China, 2008, pp. 409–414.

- D.B. Hoang, E.M. Lawrence, N. Ahmad, **V. Balasubramanian**, C.S. Homer, M. Foureur, and N. Leap, “Assistive Care Loop with Electronic Maternity Records”, in *Proceedings of the 10th IEEE International Conference on e-Health Networking, Applications and Services*, Singapore, 2008, pp. 118–123.
- **V. Balasubramanian** and S. Murugesan, “Adaptation of Web Pages for Hand-held Devices,” in *Proceedings of the 5th International Conference on Web Engineering*, LNCS Springer, Sydney, Australia, 2005, pp. 435–440.
- S. Murugesan and **V. Balasubramanian**, “Addressing the Challenges of Web Applications on Mobile Handheld Devices,” in *Proceedings of the 4th International Conference on Mobile Business*, IEEE Computer Society, Sydney, Australia, 2005, pp. 199–205.

Abstract

The exponential growth in wireless communication and micro-electro-mechanical systems gave birth to low power, low cost, multifunctional and miniature wireless sensors. Wireless sensor networks (WSNs) are finding applications in many areas, particularly in healthcare, a WSN can be used highly effectively in the form of a Body Area Wireless Sensor Network (BAWSN), to enhance the quality of the contemporary healthcare services. This tendency triggered the in-house healthcare monitoring application (HMA) with the BAWSN as their monitoring component. The quality of life of the patients can be improved with the deployment of a BAWSN because the patients and the care staff would be able to monitor the health condition, to access the electronic medical records and communicate with each other by using a Personal Digital Assistant (PDA) or smart phone in their hands, regardless of where they are located or what their duties might be. This remote monitoring and consultations might reduce the traditional stressful and costly exercise of frequent hospitalisation.

The rising costs of healthcare in many developed countries have influenced the introduction of the HMA into existing healthcare practices. However, the successful deployment of the civilian and commercial HMA is still in its infancy stages for some of the following reasons: (a) a generic framework for building the HMA is needed for the application developer to materialise the concepts, (b) an effective set of measurable properties is required for the users to evaluate the performance of the BAWSN used in the HMA, (c) the method used for measuring the performance of a monitoring application is not suitable for the BAWSN used in the HMA, (d) any unauthorized modifications to health data may induce a false diagnosis and put a life in danger; this must be considered together with the privacy of the users in the application, without any compromise and, (e) the body sensors operate in close proximity to the human body, therefore it is imperative to ensure the operational safety of BAWSN.

Our research shows that the dependable properties of an application encompass most of the essential demands for the actual realization of the HMA. It also shows

that the dependable system can deliver services that can be justifiably trusted. The focus of this work is on the BAWSN, because the service ability of the HMA depends heavily on its monitoring component, the BAWSN. Therefore, in this thesis, we define and investigate six fundamental effective properties, maintainability, reliability, availability, confidentiality, integrity and safety, based on BAWSN requirements under the dependability framework. We built an Assistive Care Loop Framework (ACLF), a maintainable BAWSN-based HMA, for monitoring pregnant women based on an event-driven architecture that follows layered functionalities. A novel critical time parameter is introduced to satisfy the needs of the medical diagnoses that require the BAWSN to rely on the collection of data within a critical time from all of the source sensor nodes. The process and issues involved in the medical diagnoses are also explored; these influenced the characterization of the BAWSN as a time critical application. Based on the critical time, we derived non-device-centric time-specific parameters with which we formulated a novel empirical measure model for the transmission reliability and availability in the BAWSN. To compare the empirical measure model, analogous theoretical models are also derived for the BAWSN's reliability and availability. To safeguard the patient's health data and also to ensure their privacy in the BAWSN, we proposed a secure adaptive triple-key scheme (α TKS) in such a way that it incurs minimal overheads while meeting the stringent time requirements of the BAWSN.

The deployed ACLF exemplifies the concept of the HMA being used not only for alerting the care staff in case of emergency, but also for analysing a patient's progress over a period of time to *forecast* any emergency situation. The performance of the proposed measure models and the implemented α TKS from our real-time test-bed would enable the application developer to measure the performance of the BAWSN-based application because, to the best of our knowledge, no measure exists in the literature as such for a BAWSN-based application. The formulated minimal set of dependable properties might serve as a *benchmark* for evaluating the performance of a BAWSN, and in turn, the HMA.

Table of Contents

UNIVERSITY OF TECHNOLOGY, SYDNEY	i
Acknowledgement	ix
List of Award and Publications	xiii
Abstract	xv
List of Figures	xxiii
List of Tables	xxv
1 Introduction	1
1.1 Background	1
1.2 Problem statement	4
1.3 Motivation	7
1.4 Contribution	8
1.4.1 Time-critical applications	10
1.4.2 Time-specific parameters	10
1.4.3 Measure model for Reliability and Availability	11
1.4.4 Security scheme for the BAWSN	11
1.4.5 A generic Healthcare Monitoring Framework	12
1.4.6 Challenges in Mobile Application Development	12
1.5 Significance	13
1.6 Approach	15
1.7 Organization of the thesis	17
2 Related Work	19
2.1 Introduction	19
2.2 The BAWSN	20
2.2.1 Components of a BAWSN	21
2.2.2 Communication standards for the BAWSN	22
2.2.3 Types of body sensors	24
2.2.4 Mote	26
2.3 Healthcare monitoring application	27
2.4 Medical diagnosis	30
2.5 Application system properties	33

2.6	Reliability.....	35
2.7	Availability.....	37
2.8	Security.....	38
2.9	Safety	38
2.10	Test-bed development.....	41
2.10.1	Presentation of web pages in mobile	41
2.11	Evaluation of related works	45
2.12	Conclusion.....	46
3	Healthcare Applications	47
3.1	Introduction.....	47
3.2	Assistive Maternity Care Application	48
3.2.1	AMC Architecture	50
3.3	Electronic Medical Record.....	52
3.3.1	Active Database System	53
3.4	Mobile Web Application	53
3.4.1	Mobile web page adaptation.....	53
3.4.2	Design aspects in mobile web pages.....	55
3.4.3	Obi-MATE.....	56
3.4.4	Security in Mobile Web Applications	59
3.4.5	Usability Study of AMC application.....	59
3.5	Conclusion	60
4	Dependability of Body Area Wireless Sensor Network.....	61
4.1	Introduction.....	61
4.2	BAWSN Architecture.....	62
4.3	Justification for dependable properties	65
4.4	Definition of dependability.....	66
4.4.1	Properties of dependability	67
4.4.2	Dependable properties description for a BAWSN	68
4.5	Use of dependable properties in fault management	70
4.6	Conclusion	72
5	A Maintainable Healthcare Framework	73
5.1	Introduction.....	73
5.2	Architecture-level maintainability.....	74

5.3	Event-driven Paradigm.....	75
5.3.1	Event-driven architecture.....	76
5.3.2	Event Flow Layers.....	76
5.4	ACLF Architecture	78
5.4.1	Functioning of ACLF	79
5.4.2	ACLF Layer Functions.....	79
5.5	Implementation of the ACLF.....	83
5.5.1	BAWSN	84
5.5.2	Health Server System.....	88
5.5.3	Mode of Monitoring.....	91
5.6	Conclusion	93
6	Time Critical Parameters	95
6.1	Introduction.....	95
6.2	Time-critical application.....	96
6.3	Selected Parameters	100
6.3.1	Critical time rate	101
6.3.2	Derivation of Parameters.....	102
6.4	Conclusion	108
7	Availability and Reliability Measure Model.....	109
7.1	Introduction.....	109
7.2	Availability Measure Model	110
7.2.1	Availability mathematical model.....	112
7.3	Reliability measure model.....	118
7.3.1	Reliability mathematical model.....	121
7.4	Critical time measurement.....	123
7.4.1	Availability and Reliability measurement.....	126
7.5	Conclusion	138
8	Security and Safety.....	139
8.1	Introduction.....	139
8.2	Security.....	141
8.2.1	A Secure Triple Key Scheme	142
8.2.2	Security Requirements for the BAWSN.....	144
8.2.3	Adaptive Triple Key Scheme	144
8.3	Safety.....	149

8.3.1	Challenges and Approach for BAWSN Safety.....	149
8.3.2	Safety Standards for Medical Devices.....	154
8.4	Conclusion	156
9	Performance Evaluation of the BAWSN.....	157
9.1	Introduction.....	157
9.2	Experimental facilities	159
9.3	Experimental Results.....	161
9.3.1	Performance of security scheme.....	165
9.4	Simulation Results	167
9.4.1	Reliability probabilistic model.....	167
9.4.2	Availability stochastic model.....	169
9.4.3	Battery discharge in a secured BAWSN	170
9.5	Achieving maintainability	172
9.5.1	Healthcare Support System.....	173
9.6	Conclusion	176
10	Conclusion and Future Work.....	177
10.1	Introduction	177
10.2	Summary of research outcomes	178
10.3	Future work.....	181
A	HMA Implementation Manual.....	183
A.1	Setting up <i>Ad Hoc</i> Network Connections	184
A.1.1	Server Network Setup	185
A.1.2	Client Network Setup.....	185
A.1.3	Establishing the <i>ad hoc</i> connection.....	186
A.1.4	Testing the connectivity	186
A.2	Setting up the Web Server	187
A.2.1	Installing Zend Server (Community Edition).....	187
A.2.2	Setting up the httpd.conf file	188
A.2.3	Setting up the php.ini file	189
A.2.4	Testing the Server's Name	189
A.2.5	Calling Java Class from PHP	189
A.3	Setting up the Website.....	191
A.3.1	Web Application Setup Steps	191

A.3.2	Database Setup Steps	192
A.4	Using the Website	192
A.4.1	Steps to use administrator's module	192
A.4.2	Steps to use doctor's module	193
A.4.3	Steps to use patient's module	193
A.4.4	Accepting the patient's request	194
A.5	Web service components	194
A.6	Sending SMS using Third Party Software.....	195
	GSM Modem driver installation.....	195
	Configuring the Apache2.2-Zend Service	195
B	HMA API and Code Snippets	197
B.1	API Details	197
B.2	Code Snippets.....	206
C	Nomenclature.....	217
References	221

List of Figures

Figure 2-1 Different Healthcare Monitoring Applications (a) HipGuard pants for hip patient rehabilitation [103] (b) Wireless Mica 2 based pulse oximeter sensor [102] (c) Smart LifeShirt [98] (d) An ECG module communicating with the base station [100]	28
Figure 3-1 Assistive Maternity Care Application	51
Figure 3-2: Presentation of web pages on hand-held devices	54
Figure 3-3 Desktop Interface for AMC application (a) details of visit, (b) list of options for viewing	58
Figure 3-4 PDA Interfaces for AMC application, (a) details of visits, (b) drop-down menu for viewing selection, (c) woman's details	58
Figure 4-1 Different Domains of an Healthcare Monitoring Application	63
Figure 4-2 Properties of Dependability	67
Figure 5-1 Event Processing Flows in Event Driven Architecture	77
Figure 5-2 (a) ED Architecture for HMA (b) ACLF Layers Framed Based on ED Architecture	81
Figure 5-3 Assistive Care Loop Framework	84
Figure 5-4 Implemented Hardware Components in BAWSN	85
Figure 5-5 The GUI for the Obi-MATE 2 (a) continuous monitoring, (b) to get current data	87
Figure 5-6 Alert Message using (a) notification email (b) an SMS	90
Figure 5-7 Behaviour of the AMC application when receiving data from a PDA, (a) monitored sensor data (b) user input data using Web Page	92
Figure 6-1 Packet timing from sensors at PDA showing error and lost packets in BAWSN	102
Figure 6-2 Packet timing from sensors at PDA showing delayed packets in BAWSN	106
Figure 7-1 Packet timing from sensors at PDA showing overlap packets for availability measure	111
Figure 7-2 State transition model, (a) for WSN cluster, (b) for BAWSN	113
Figure 7-3 A comparison of a (a) WSN cluster and (b) the BAWSN architecture	116
Figure 7-4 A two-state Markov chain for a BAWSN with the State probabilities	118
Figure 7-5 Packet timing from sensors at PDA showing overlap packets for reliability measure	119
Figure 7-6. (a) A Body Area Wireless Sensor Network, (b) the corresponding graph model	121
Figure 7-7 Evaluation of critical time in BAWSN (a) IDEAL CASE (b) NORMAL CASE (c) SPECIAL CASE (d) RESTART CASE	124
Figure 7-8 Packet timing of sensors in the BAWSN for IDEAL case	127
Figure 7-9 Packet timing of the sensors in the BAWSN for the NORMAL CASE	133
Figure 8-1 Implementation of α TKS in the BAWSN components	147
Figure 8-2 Model based engineering methodology [177]	152
Figure 9-1 Experimental setup for the BAWSN	160

Figure 9-2 Performance of a BAWSN (a) Operational Reliability (b) Operational Availability (c) Operational Throughput (d) number of error and lost packets during operational time	163
Figure 9-3 Performance of a BAWSN with and without α TKS, (a) Operational Reliability (b) Packet Error	166
Figure 9-4 Effect of probability in base station buffer on reliability of a BAWSN: (a) ρ_{ERR} vs ρ_E or ρ_B . (b) ρ_{ERR} vs ρ_B for different values of ρ_E . (c) ρ_{UR} vs ρ_B for different values of ρ_{LACK} and negligible ρ_E . (d) ρ_{UR} vs ρ_B or ρ_E for different values of ρ_{LACK} (e) ρ_{UR} vs ρ_{LACK} for different values of ρ_E . (f) ρ_{ERR} vs ρ_E for different values of ρ_B	169
Figure 9-5 Healthy state probability transition for the BAWSN	170
Figure 9-6 Performance of a BAWSN with and without an α TKS, (a) Battery Discharge (b) Throughput	171
Figure 9-7 Healthcare Support System (a) Hardware Setup (b) Doctor's view of patients information (c) Doctors continuous monitoring of the patient	175
Figure B-1 Console output of AT commands when sending an SMS	216

List of Tables

Table 2-1 Specification of Three Different Types of Sensors26

Table 2-2 Comparison of Web Page Presentation Methods44

Table 6-1 Three sets of health data received at sequential time instant (CASE 1)98

Table 6-2 Three sets of health data received at sequential time instant (CASE 2)99

Table 7-1 Rate and probabilistic parameters for availability in BAWSN 114

Table 8-1 Comparison of overheads in TinySec, MiniSec and STKS 143

Table A-1 List of Software and Hardware used in HMA..... 184

Table B-1 HealthCare Support System API descibtion..... 197

Table B-2 Obi MATE 2 API Description..... 204

